


# BC-NRBC

## HEMATOLOGY CONTROLS

<b>CONTROL</b>
----------------

### ASSAY VALUES AND EXPECTED RANGES

<b>LOT</b>	<b>MN0318</b>
	<b>2018-05-10</b>

Instrument	Parameter	Low	Normal	High
		<div>LOT</div> MN0318-1	<div>LOT</div> MN0318-2	<div>LOT</div> MN0318-3
BC-6800, BC-6600	WBC_N x 10 <sup>9</sup> /L *	9.30 ± 1.00	9.20 ± 1.00	3.30 ± 0.50
	NRBC x 10 <sup>9</sup> /L	0.020 ± 0.020	0.632 ± 0.192	0.626 ± 0.150
	NRBC %/100WBC	0.20 ± 0.20	6.60 ± 2.00	16.65 ± 4.00

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.



**Mindray**

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China

Tel: +86 755 81888998

Fax: +86 755 26582680

<b>EC</b>
-----------

<b>REP</b>
------------

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726