


BC-NRBC

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT	MN0518
	2018-07-10

Instrument	Parameter	Low	Normal	High
		<div>LOT</div> MN0518-1	<div>LOT</div> MN0518-2	<div>LOT</div> MN0518-3
BC-6800, BC-6600	WBC_N x 10 ⁹ /L *	9.10 ± 1.00	9.15 ± 1.00	3.35 ± 0.50
QC Mode	NRBC x 10 ⁹ /L	0.020 ± 0.020	0.518 ± 0.188	0.655 ± 0.152
	NRBC %/100WBC	0.20 ± 0.20	5.50 ± 2.00	17.20 ± 4.00

* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.



Mindray

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China

Tel: +86 755 81888998

Fax: +86 755 26582680

EC

REP

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726