

Research use only

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT ME0519A

 2019-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME0519A-1	LOT	ME0519A-2	LOT	ME0519A-3
MINDRAY BC-6800, BC-6600	IPF %	12.5	± 3.5	12.5	± 3.5	11.0	± 3.5
	PLT-O 10 ⁹ /L	129	± 20	293	± 40	538	± 60
QC Mode							

Instrument	Parameter	Low		Normal		High	
		LOT	ME0519A-1	LOT	ME0519A-2	LOT	ME0519A-3
MINDRAY BC-6200, BC-6000Plus, BC-6100Plus	IPF %	4.7	± 15.0	4.6	± 15.0	4.7	± 15.0
	RBC-O 10 ¹² /L	5.79	± 0.35	4.71	± 0.30	2.23	± 0.20
	PLT-O 10 ⁹ /L	128	± 20	296	± 40	537	± 60
QC Mode							

Instrument	Parameter	Low		Normal		High	
		LOT	ME0519A-1	LOT	ME0519A-2	LOT	ME0519A-3
MINDRAY BC-6800Plus, BC-6700Plus, BC-6600Plus	IPF %	4.4	± 15.0	4.6	± 15.0	4.3	± 15.0
	RBC-O 10 ¹² /L	5.62	± 0.35	4.63	± 0.30	2.26	± 0.20
	PLT-O 10 ⁹ /L	125	± 20	298	± 40	554	± 60
QC Mode							

Research use only