

Research use only

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT ME0719A

 **2019-09-10**

Instrument	Parameter	Low		Normal		High	
		LOT	ME0719A-1	LOT	ME0719A-2	LOT	ME0719A-3
MINDRAY BC-6800, BC-6600	PLT-O $10^9/L$	131	\pm 20	290	\pm 40	528	\pm 60
QC Mode							

Instrument	Parameter	Low		Normal		High	
		LOT	ME0719A-1	LOT	ME0719A-2	LOT	ME0719A-3
MINDRAY BC-6200, BC-6000Plus, BC-6100Plus	RBC-O $10^{12}/L$	5.76	\pm 0.35	4.65	\pm 0.30	2.29	\pm 0.20
QC Mode							

Instrument	Parameter	Low		Normal		High	
		LOT	ME0719A-1	LOT	ME0719A-2	LOT	ME0719A-3
MINDRAY BC-6800Plus, BC-6700Plus, BC-6600Plus	RBC-O $10^{12}/L$	5.63	\pm 0.35	4.57	\pm 0.30	2.33	\pm 0.20
QC Mode							

Research use only