

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT ME0719A

 **2019-09-10**

Instrument	Parameter	Low		Normal		High	
		LOT	ME0719A-1	LOT	ME0719A-2	LOT	ME0719A-3
MINDRAY	Ret# $10^{12}/L$	0.0667	± 0.0536	0.1979	± 0.0792	0.1771	± 0.0531
BC-6800, BC-6600	Ret% %	1.17	± 0.94	4.35	± 1.74	7.60	± 2.28
	IRF %	7.6	± 15.0	7.5	± 13.0	6.3	± 25.0
QC Mode	LFR %	92.4	± 15.0	92.5	± 10.0	93.7	± 25.0
	MFR %	6.3	± 12.0	6.2	± 10.0	4.8	± 22.0
	HFR %	1.3	± 3.0	1.3	± 3.0	1.5	± 3.0
	RHE pg	24.5	± 10.0	24.0	± 10.0	21.5	± 10.0
	IPF %	13.5	± 3.5	11.0	± 3.5	13.0	± 3.5
	RBC-O* $10^{12}/L$	5.51	± 0.35	4.47	± 0.30	2.24	± 0.20

Before using, refer to the instruction sheet for mixing directions.

*Research use only.

Instrument	Parameter	Low		Normal		High	
		LOT	ME0719A-1	LOT	ME0719A-2	LOT	ME0719A-3
MINDRAY	Ret# $10^{12}/L$	0.0782	± 0.0230	0.2111	± 0.0990	0.1835	± 0.0770
BC-6200,	Ret% %	1.33	± 0.40	4.50	± 2.00	7.71	± 3.00
BC-6000Plus,	IRF %	12.1	± 15.0	11.2	± 13.0	9.6	± 25.0
BC-6100Plus	LFR %	87.9	± 15.0	88.8	± 10.0	90.4	± 25.0
	MFR %	9.2	± 12.0	8.6	± 10.0	5.5	± 22.0
QC Mode	HFR %	2.9	± 3.0	2.6	± 3.0	4.1	± 3.0
	RHE pg	29.0	± 10.0	29.1	± 10.0	24.4	± 10.0
	IPF %	4.5	± 15.0	4.5	± 15.0	4.5	± 15.0

Before using, refer to the instruction sheet for mixing directions.

Instrument	Parameter	Low		Normal		High	
		LOT	ME0719A-1	LOT	ME0719A-2	LOT	ME0719A-3
MINDRAY	Ret# $10^{12}/L$	0.0892	± 0.0230	0.2121	± 0.0990	0.1804	± 0.0770
BC-6800Plus,	Ret% %	1.56	± 0.40	4.59	± 2.00	7.61	± 3.00
BC-6700Plus,	IRF %	10.7	± 15.0	10.8	± 13.0	9.7	± 25.0
BC-6600Plus	LFR %	89.3	± 15.0	89.2	± 10.0	90.3	± 25.0
	MFR %	8.2	± 12.0	8.0	± 10.0	5.3	± 22.0
QC Mode	HFR %	2.5	± 3.0	2.8	± 3.0	4.4	± 3.0
	RHE pg	28.9	± 10.0	28.7	± 10.0	23.9	± 10.0
	IPF %	4.4	± 15.0	4.3	± 15.0	4.5	± 15.0

Before using, refer to the instruction sheet for mixing directions.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, ShenZhen 518057, P.R.Ch

Tel: +86 755 26582479 26582888

Fax: +86 755 26582934 26582500

EC REP

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726

