

S30 CALIBRATOR

FOR MINDRAY HEMATOLOGY ANALYZERS

LOT

BC10119C-1



2019-05-19

Calibration Values

Analyzer	Parameter	Values	Range
BC-3600, BC-3300	WBC $\times 10^9/L$	9.1	± 0.3
	RBC $\times 10^{12}/L$	4.39	± 0.10
	HGB g/L	133	± 2
	MCV fL	87.8	± 2.0
	PLT $\times 10^9/L$	288	± 15
BC-3200, BC-3000CT	WBC $\times 10^9/L$	9.0	± 0.3
	RBC $\times 10^{12}/L$	4.41	± 0.10
	HGB g/L	134	± 2
	MCV fL	88.4	± 2.0
	PLT $\times 10^9/L$	263	± 15
BC-2900, BC-1800, BC-3000 Plus	WBC $\times 10^9/L$	9.1	± 0.3
	RBC $\times 10^{12}/L$	4.44	± 0.10
	HGB g/L	135	± 2
	MCV fL	88.5	± 2.0
	PLT $\times 10^9/L$	265	± 15
BC-2800, BC-2800Vet, BC-2600, BC-2600Vet	WBC $\times 10^9/L$	9.1	± 0.3
	RBC $\times 10^{12}/L$	4.31	± 0.10
	HGB g/L	135	± 2
	MCV fL	89.0	± 2.0
	PLT $\times 10^9/L$	275	± 15
BC-2300, BC-2100	WBC $\times 10^9/L$	9.0	± 0.3
	RBC $\times 10^{12}/L$	4.11	± 0.10
	HGB g/L	133	± 2
	MCV fL	87.7	± 2.0
	PLT $\times 10^9/L$	289	± 15
BC-20s, BC-21s BC-30s, BC-31s BC-30, BC-31, BC-30e	WBC $\times 10^9/L$	9.2	± 0.3
	RBC $\times 10^{12}/L$	4.27	± 0.10
	HGB g/L	136	± 2
	MCV fL	88.7	± 2.0
	PLT $\times 10^9/L$	266	± 15
BC-20, BC-21 BC-10, BC-11, BC-10e	WBC $\times 10^9/L$	9.1	± 0.3
	RBC $\times 10^{12}/L$	4.26	± 0.10
	HGB g/L	136	± 2
	MCV fL	89.3	± 2.0
	PLT $\times 10^9/L$	280	± 15