

S30 CALIBRATOR

FOR MINDRAY HEMATOLOGY ANALYZERS

LOT

BC10319C-1



2019-07-20

Calibration Values

Analyzer	Parameter	Values	Range
BC-3600, BC-3300	WBC $\times 10^9/L$	8.8	± 0.3
	RBC $\times 10^{12}/L$	4.32	± 0.10
	HGB g/L	134	± 2
	MCV fL	89.5	± 2.0
	PLT $\times 10^9/L$	257	± 15
BC-3200, BC-3000CT	WBC $\times 10^9/L$	8.8	± 0.3
	RBC $\times 10^{12}/L$	4.36	± 0.10
	HGB g/L	135	± 2
	MCV fL	89.9	± 2.0
	PLT $\times 10^9/L$	246	± 15
BC-2900, BC-1800, BC-3000 Plus	WBC $\times 10^9/L$	8.8	± 0.3
	RBC $\times 10^{12}/L$	4.36	± 0.10
	HGB g/L	136	± 2
	MCV fL	90.1	± 2.0
	PLT $\times 10^9/L$	247	± 15
BC-2800, BC-2800Vet, BC-2600, BC-2600Vet	WBC $\times 10^9/L$	8.8	± 0.3
	RBC $\times 10^{12}/L$	4.29	± 0.10
	HGB g/L	135	± 2
	MCV fL	91.1	± 2.0
	PLT $\times 10^9/L$	253	± 15
BC-2300, BC-2100	WBC $\times 10^9/L$	8.8	± 0.3
	RBC $\times 10^{12}/L$	4.10	± 0.10
	HGB g/L	133	± 2
	MCV fL	89.4	± 2.0
	PLT $\times 10^9/L$	262	± 15
BC-20s, BC-21s BC-30s, BC-31s BC-30, BC-31, BC-30e	WBC $\times 10^9/L$	9.0	± 0.3
	RBC $\times 10^{12}/L$	4.16	± 0.10
	HGB g/L	136	± 2
	MCV fL	90.2	± 2.0
	PLT $\times 10^9/L$	245	± 15
BC-20, BC-21 BC-10, BC-11, BC-10e	WBC $\times 10^9/L$	8.8	± 0.3
	RBC $\times 10^{12}/L$	4.21	± 0.10
	HGB g/L	135	± 2
	MCV fL	91.0	± 2.0
	PLT $\times 10^9/L$	258	± 15