

# BC-5D

## HEMATOLOGY CONTROLS

<b>CONTROL</b>
----------------

ASSAY VALUES AND EXPECTED RANGES

<b>LOT</b>
------------

**BC1901B****2019-03-10**

Instrument	Parameter	Low		Normal		High		+
		LOT	BC1901BL	LOT	BC1901BN	LOT	BC1901BH	
<b>BC-5800,BC-5600</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.50	$\pm 0.50$	8.05	$\pm 1.00$	18.40	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.65	$\pm 0.32$	4.35	$\pm 0.65$	11.59	$\pm 1.48$	
	Lym# $\times 10^9/L$	1.38	$\pm 0.32$	2.50	$\pm 0.65$	3.86	$\pm 1.48$	
	Mon# $\times 10^9/L$	0.26	$\pm 0.21$	0.56	$\pm 0.48$	1.29	$\pm 1.11$	
	Eos# $\times 10^9/L$	0.18	$\pm 0.15$	0.56	$\pm 0.48$	1.47	$\pm 1.29$	
	Bas# $\times 10^9/L$	0.04	$\pm 0.04$	0.08	$\pm 0.08$	0.18	$\pm 0.18$	
	Neu%	47.0	$\pm 9.0$	54.0	$\pm 8.0$	63.0	$\pm 8.0$	
	Lym%	39.5	$\pm 9.0$	31.0	$\pm 8.0$	21.0	$\pm 8.0$	
	Mon%	7.5	$\pm 6.0$	7.0	$\pm 6.0$	7.0	$\pm 6.0$	
	Eos%	5.0	$\pm 4.0$	7.0	$\pm 6.0$	8.0	$\pm 7.0$	
	Bas%	1.0	$\pm 1.0$	1.0	$\pm 1.0$	1.0	$\pm 1.0$	
	RBC $\times 10^{12}/L$	2.09	$\pm 0.18$	4.23	$\pm 0.24$	5.06	$\pm 0.30$	
	HGB g/L	59	$\pm 4$	134	$\pm 6$	171	$\pm 8$	
	HCT %	18.0	$\pm 1.5$	41.0	$\pm 2.0$	53.1	$\pm 2.4$	
	MCV fL	86.0	$\pm 5.0$	97.0	$\pm 5.0$	105.0	$\pm 5.0$	
	MCH pg	28.2	$\pm 2.5$	31.7	$\pm 2.5$	33.8	$\pm 2.5$	
	MCHC g/L	328	$\pm 30$	327	$\pm 30$	322	$\pm 30$	
	RDW-CV %	14.0	$\pm 3.0$	13.5	$\pm 3.0$	12.8	$\pm 3.0$	
	RDW-SD fL	42.0	$\pm 8.0$	45.0	$\pm 8.0$	46.0	$\pm 8.0$	
	PLT $\times 10^9/L$	57	$\pm 20$	253	$\pm 40$	500	$\pm 60$	
	MPV fL	8.1	$\pm 3.0$	8.1	$\pm 3.0$	7.9	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.210	$\pm 0.100$	0.400	$\pm 0.200$	
	PDW*	18.4	$\pm 3.0$	16.8	$\pm 3.0$	16.0	$\pm 3.0$	
	P-LCC $\times 10^9/L$	15	$\pm 15$	58	$\pm 25$	95	$\pm 35$	
	P-LCR %	25.0	$\pm 10.0$	21.5	$\pm 10.0$	18.0	$\pm 10.0$	
<b>BC-5500,BC-5200</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.40	$\pm 0.50$	7.80	$\pm 1.00$	17.80	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.58	$\pm 0.31$	4.29	$\pm 0.63$	10.95	$\pm 1.43$	
	Lym# $\times 10^9/L$	1.36	$\pm 0.31$	2.34	$\pm 0.63$	3.74	$\pm 1.43$	
	Mon# $\times 10^9/L$	0.22	$\pm 0.21$	0.47	$\pm 0.40$	1.25	$\pm 1.08$	
	Eos# $\times 10^9/L$	0.20	$\pm 0.17$	0.62	$\pm 0.55$	1.69	$\pm 1.43$	
	Bas# $\times 10^9/L$	0.03	$\pm 0.03$	0.08	$\pm 0.08$	0.18	$\pm 0.18$	
	Neu%	46.5	$\pm 9.0$	55.0	$\pm 8.0$	61.5	$\pm 8.0$	
	Lym%	40.0	$\pm 9.0$	30.0	$\pm 8.0$	21.0	$\pm 8.0$	
	Mon%	6.5	$\pm 6.0$	6.0	$\pm 5.0$	7.0	$\pm 6.0$	
	Eos%	6.0	$\pm 5.0$	8.0	$\pm 7.0$	9.5	$\pm 8.0$	
	Bas%	1.0	$\pm 1.0$	1.0	$\pm 1.0$	1.0	$\pm 1.0$	
	RBC $\times 10^{12}/L$	2.12	$\pm 0.18$	4.23	$\pm 0.24$	5.02	$\pm 0.30$	
	HGB g/L	63	$\pm 4$	143	$\pm 6$	182	$\pm 8$	
	HCT %	17.5	$\pm 1.5$	39.1	$\pm 2.0$	49.9	$\pm 2.4$	
	MCV fL	82.5	$\pm 5.0$	92.5	$\pm 5.0$	99.5	$\pm 5.0$	
	MCH pg	29.7	$\pm 2.5$	33.8	$\pm 2.5$	36.3	$\pm 2.5$	
	MCHC g/L	360	$\pm 30$	365	$\pm 30$	364	$\pm 30$	
	RDW-CV %	11.4	$\pm 3.0$	11.0	$\pm 3.0$	10.5	$\pm 3.0$	
	RDW-SD fL	29.5	$\pm 8.0$	33.0	$\pm 8.0$	35.0	$\pm 8.0$	
	PLT $\times 10^9/L$	55	$\pm 20$	247	$\pm 40$	482	$\pm 60$	
	MPV fL	9.0	$\pm 3.0$	8.9	$\pm 3.0$	8.6	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.220	$\pm 0.100$	0.415	$\pm 0.200$	
	PDW*	17.6	$\pm 3.0$	16.9	$\pm 3.0$	16.2	$\pm 3.0$	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

**BC1901B**  
**2019-03-10**

Instrument	Parameter	Low		Normal		High		++
		LOT	BC1901BL	LOT	BC1901BN	LOT	BC1901BH	
<b>BC-5390</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.35	$\pm 0.50$	7.85	$\pm 1.00$	17.60	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.68	$\pm 0.31$	4.44	$\pm 0.63$	11.35	$\pm 1.59$	
	Lym# $\times 10^9/L$	1.22	$\pm 0.31$	2.24	$\pm 0.63$	3.26	$\pm 1.24$	
	Mon# $\times 10^9/L$	0.27	$\pm 0.21$	0.55	$\pm 0.48$	1.23	$\pm 1.06$	
	Eos# $\times 10^9/L$	0.18	$\pm 0.17$	0.63	$\pm 0.48$	1.76	$\pm 1.41$	
	Bas# $\times 10^9/L$	0.84	$\pm 0.34$	2.13	$\pm 0.79$	5.35	$\pm 1.76$	
	Neu%	50.0	$\pm 9.0$	56.5	$\pm 8.0$	64.5	$\pm 9.0$	
	Lym%	36.5	$\pm 9.0$	28.5	$\pm 8.0$	18.5	$\pm 7.0$	
	Mon%	8.0	$\pm 6.0$	7.0	$\pm 6.0$	7.0	$\pm 6.0$	
	Eos%	5.5	$\pm 5.0$	8.0	$\pm 6.0$	10.0	$\pm 8.0$	
	Bas%	25.2	$\pm 10.0$	27.1	$\pm 10.0$	30.4	$\pm 10.0$	
	RBC $\times 10^{12}/L$	2.02	$\pm 0.18$	4.18	$\pm 0.24$	5.01	$\pm 0.30$	
	HGB g/L	55	$\pm 4$	126	$\pm 6$	161	$\pm 8$	
	HCT %	17.2	$\pm 1.5$	39.3	$\pm 2.0$	51.1	$\pm 2.4$	
	MCV fL	85.0	$\pm 5.0$	94.0	$\pm 5.0$	102.0	$\pm 5.0$	
	MCH pg	27.2	$\pm 2.5$	30.1	$\pm 2.5$	32.1	$\pm 2.5$	
	MCHC g/L	320	$\pm 30$	321	$\pm 30$	315	$\pm 30$	
	RDW-CV %	13.6	$\pm 3.0$	13.5	$\pm 3.0$	12.9	$\pm 3.0$	
	RDW-SD fL	45.5	$\pm 8.0$	48.0	$\pm 8.0$	49.5	$\pm 8.0$	
	PLT $\times 10^9/L$	55	$\pm 20$	239	$\pm 40$	476	$\pm 60$	
	MPV fL	11.4	$\pm 3.0$	11.0	$\pm 3.0$	10.6	$\pm 3.0$	
<b>BC-5390 CRP</b> <b>QC Mode</b>	WBC $\times 10^9/L$	3.35	$\pm 0.50$	7.80	$\pm 1.00$	18.00	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.66	$\pm 0.31$	4.41	$\pm 0.63$	11.52	$\pm 1.62$	
	Lym# $\times 10^9/L$	1.24	$\pm 0.31$	2.22	$\pm 0.63$	3.51	$\pm 1.26$	
	Mon# $\times 10^9/L$	0.27	$\pm 0.21$	0.51	$\pm 0.48$	1.17	$\pm 1.08$	
	Eos# $\times 10^9/L$	0.18	$\pm 0.17$	0.66	$\pm 0.55$	1.80	$\pm 1.44$	
	Bas# $\times 10^9/L$	0.85	$\pm 0.34$	2.11	$\pm 0.79$	5.45	$\pm 1.80$	
	Neu%	49.5	$\pm 9.0$	56.5	$\pm 8.0$	64.0	$\pm 9.0$	
	Lym%	37.0	$\pm 9.0$	28.5	$\pm 8.0$	19.5	$\pm 7.0$	
	Mon%	8.0	$\pm 6.0$	6.5	$\pm 6.0$	6.5	$\pm 6.0$	
	Eos%	5.5	$\pm 5.0$	8.5	$\pm 7.0$	10.0	$\pm 8.0$	
	Bas%	25.3	$\pm 10.0$	27.0	$\pm 10.0$	30.3	$\pm 10.0$	
	RBC $\times 10^{12}/L$	2.01	$\pm 0.18$	4.16	$\pm 0.24$	5.00	$\pm 0.30$	
	HGB g/L	55	$\pm 4$	126	$\pm 6$	161	$\pm 8$	
	HCT %	16.8	$\pm 1.5$	39.5	$\pm 2.0$	51.5	$\pm 2.4$	
	MCV fL	83.5	$\pm 5.0$	95.0	$\pm 5.0$	103.0	$\pm 5.0$	
	MCH pg	27.4	$\pm 2.5$	30.3	$\pm 2.5$	32.2	$\pm 2.5$	
	MCHC g/L	328	$\pm 30$	319	$\pm 30$	313	$\pm 30$	
	RDW-CV %	14.5	$\pm 3.0$	13.5	$\pm 3.0$	13.0	$\pm 3.0$	
	RDW-SD fL	43.0	$\pm 8.0$	46.0	$\pm 8.0$	48.0	$\pm 8.0$	
	PLT $\times 10^9/L$	48	$\pm 20$	235	$\pm 40$	475	$\pm 60$	
	MPV fL	9.2	$\pm 3.0$	8.5	$\pm 3.0$	8.0	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.200	$\pm 0.100$	0.380	$\pm 0.200$	
	PDW*	16.2	$\pm 3.0$	16.4	$\pm 3.0$	15.9	$\pm 3.0$	
	P-LCC $\times 10^9/L$	15	$\pm 15$	46	$\pm 25$	76	$\pm 35$	
	P-LCR %	23.5	$\pm 10.0$	19.5	$\pm 10.0$	16.0	$\pm 10.0$	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

## ASSAY VALUES AND EXPECTED RANGES

LOT

BC1901B



2019-03-10

Instrument	Parameter	Low		Normal		High		+++
		LOT	BC1901BL	LOT	BC1901BN	LOT	BC1901BH	
BC-5300,BC-5100	WBC $\times 10^9/L$	3.25	$\pm 0.50$	7.80	$\pm 1.00$	17.80	$\pm 2.50$	
BC-5380,BC-5180	Neu# $\times 10^9/L$	1.64	$\pm 0.30$	4.56	$\pm 0.63$	11.71	$\pm 1.43$	
QC Mode	Lym# $\times 10^9/L$	1.25	$\pm 0.30$	2.22	$\pm 0.63$	3.51	$\pm 1.43$	
(Software version lower than 1.24.00.16860)	Mon# $\times 10^9/L$	0.16	$\pm 0.13$	0.31	$\pm 0.24$	0.71	$\pm 0.54$	
	Eos# $\times 10^9/L$	0.20	$\pm 0.17$	0.70	$\pm 0.55$	1.87	$\pm 1.43$	
	Bas# $\times 10^9/L$	1.77	$\pm 0.33$	4.94	$\pm 0.79$	13.42	$\pm 1.78$	
	Neu%	50.5	$\pm 9.0$	58.5	$\pm 8.0$	65.8	$\pm 8.0$	
	Lym%	38.5	$\pm 9.0$	28.5	$\pm 8.0$	19.7	$\pm 8.0$	
	Mon%	5.0	$\pm 4.0$	4.0	$\pm 3.0$	4.0	$\pm 3.0$	
	Eos%	6.0	$\pm 5.0$	9.0	$\pm 7.0$	10.5	$\pm 8.0$	
	Bas%	54.5	$\pm 10.0$	63.3	$\pm 10.0$	75.4	$\pm 10.0$	
	RBC $\times 10^{12}/L$	2.03	$\pm 0.18$	4.17	$\pm 0.24$	5.00	$\pm 0.30$	
	HGB g/L	56	$\pm 4$	126	$\pm 6$	161	$\pm 8$	
	HCT %	18.3	$\pm 1.5$	42.0	$\pm 2.0$	54.3	$\pm 2.4$	
	MCV fL	90.0	$\pm 5.0$	100.8	$\pm 5.0$	108.5	$\pm 5.0$	
	MCH pg	27.6	$\pm 2.5$	30.2	$\pm 2.5$	32.2	$\pm 2.5$	
	MCHC g/L	307	$\pm 30$	300	$\pm 30$	297	$\pm 30$	
	RDW-CV %	14.0	$\pm 3.0$	13.5	$\pm 3.0$	13.0	$\pm 3.0$	
	RDW-SD fL	56.0	$\pm 8.0$	59.0	$\pm 8.0$	60.8	$\pm 8.0$	
	PLT $\times 10^9/L$	54	$\pm 20$	236	$\pm 40$	475	$\pm 60$	
	MPV fL	9.2	$\pm 3.0$	8.6	$\pm 3.0$	8.1	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.200	$\pm 0.100$	0.385	$\pm 0.200$	
	PDW*	16.8	$\pm 3.0$	16.5	$\pm 3.0$	16.0	$\pm 3.0$	
BC-5300,BC-5100	WBC $\times 10^9/L$	3.30	$\pm 0.50$	7.90	$\pm 1.00$	17.80	$\pm 2.50$	
BC-5380,BC-5180	Neu# $\times 10^9/L$	1.67	$\pm 0.30$	4.58	$\pm 0.64$	11.57	$\pm 1.43$	
QC Mode	Lym# $\times 10^9/L$	1.25	$\pm 0.30$	2.33	$\pm 0.64$	3.56	$\pm 1.43$	
(Software version 1.24.00.16860 or higher)	Mon# $\times 10^9/L$	0.17	$\pm 0.14$	0.32	$\pm 0.24$	0.71	$\pm 0.54$	
	Eos# $\times 10^9/L$	0.21	$\pm 0.17$	0.67	$\pm 0.48$	1.96	$\pm 1.61$	
	Bas# $\times 10^9/L$	1.91	$\pm 0.34$	5.19	$\pm 0.79$	13.71	$\pm 1.79$	
	Neu%	50.5	$\pm 9.0$	58.0	$\pm 8.0$	65.0	$\pm 8.0$	
	Lym%	38.0	$\pm 9.0$	29.5	$\pm 8.0$	20.0	$\pm 8.0$	
	Mon%	5.0	$\pm 4.0$	4.0	$\pm 3.0$	4.0	$\pm 3.0$	
	Eos%	6.5	$\pm 5.0$	8.5	$\pm 6.0$	11.0	$\pm 9.0$	
	Bas%	57.8	$\pm 10.0$	65.7	$\pm 10.0$	77.0	$\pm 10.0$	
	RBC $\times 10^{12}/L$	2.02	$\pm 0.18$	4.18	$\pm 0.24$	4.97	$\pm 0.30$	
	HGB g/L	57	$\pm 4$	127	$\pm 6$	161	$\pm 8$	
	HCT %	17.7	$\pm 1.5$	41.0	$\pm 2.0$	52.7	$\pm 2.4$	
	MCV fL	87.5	$\pm 5.0$	98.0	$\pm 5.0$	106.0	$\pm 5.0$	
	MCH pg	28.2	$\pm 2.5$	30.4	$\pm 2.5$	32.4	$\pm 2.5$	
	MCHC g/L	322	$\pm 30$	310	$\pm 30$	306	$\pm 30$	
	RDW-CV %	14.0	$\pm 3.0$	13.5	$\pm 3.0$	13.0	$\pm 3.0$	
	RDW-SD fL	52.0	$\pm 8.0$	55.0	$\pm 8.0$	57.0	$\pm 8.0$	
	PLT $\times 10^9/L$	50	$\pm 20$	237	$\pm 40$	475	$\pm 60$	
	MPV fL	9.3	$\pm 3.0$	8.7	$\pm 3.0$	8.3	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.205	$\pm 0.100$	0.390	$\pm 0.200$	
	PDW*	16.4	$\pm 3.0$	16.3	$\pm 3.0$	15.8	$\pm 3.0$	

\* For Research Use Only

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.

# BC-5D

## HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

BC1901B

2019-03-10



Instrument	Parameter	Low		Normal		High		++++
		LOT	BC1901BL	LOT	BC1901BN	LOT	BC1901BH	
BC-5000,BC-5150,BC-5120 BC-5130,BC-5140,BC-5000VET QC Mode	WBC $\times 10^9/L$	3.35	$\pm 0.50$	7.80	$\pm 1.00$	17.90	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.57	$\pm 0.41$	4.36	$\pm 0.94$	11.19	$\pm 2.15$	
	Lym# $\times 10^9/L$	1.29	$\pm 0.31$	2.18	$\pm 0.63$	3.42	$\pm 1.26$	
	Mon# $\times 10^9/L$	0.28	$\pm 0.28$	0.59	$\pm 0.59$	1.34	$\pm 1.34$	
	Eos# $\times 10^9/L$	0.18	$\pm 0.18$	0.56	$\pm 0.56$	1.66	$\pm 1.66$	
	Bas# $\times 10^9/L$	0.03	$\pm 0.03$	0.10	$\pm 0.10$	0.29	$\pm 0.29$	
	Neu%	46.8	$\pm 12.0$	55.9	$\pm 12.0$	62.5	$\pm 12.0$	
	Lym%	38.6	$\pm 9.0$	28.0	$\pm 8.0$	19.1	$\pm 7.0$	
	Mon%	8.4	$\pm 8.4$	7.6	$\pm 7.6$	7.5	$\pm 7.5$	
	Eos%	5.4	$\pm 5.4$	7.2	$\pm 7.2$	9.3	$\pm 9.3$	
	Bas%	0.8	$\pm 0.8$	1.3	$\pm 1.3$	1.6	$\pm 1.6$	
	RBC $\times 10^{12}/L$	2.05	$\pm 0.18$	4.21	$\pm 0.24$	5.06	$\pm 0.30$	
	HGB g/L	56	$\pm 4$	128	$\pm 6$	166	$\pm 8$	
	HCT %	18.0	$\pm 1.5$	40.8	$\pm 2.0$	52.4	$\pm 2.4$	
	MCV fL	88.0	$\pm 5.0$	97.0	$\pm 5.0$	103.5	$\pm 5.0$	
	MCH pg	27.3	$\pm 2.5$	30.4	$\pm 2.5$	32.8	$\pm 2.5$	
	MCHC g/L	310	$\pm 30$	313	$\pm 30$	317	$\pm 30$	
	RDW-CV %	16.5	$\pm 3.0$	15.5	$\pm 3.0$	15.0	$\pm 3.0$	
	RDW-SD fL	54.0	$\pm 8.0$	56.0	$\pm 8.0$	56.5	$\pm 8.0$	
	PLT $\times 10^9/L$	50	$\pm 20$	243	$\pm 40$	504	$\pm 60$	
	MPV fL	11.3	$\pm 3.0$	10.2	$\pm 3.0$	9.6	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.250	$\pm 0.100$	0.480	$\pm 0.200$	
	PDW*	15.9	$\pm 3.0$	16.4	$\pm 3.0$	16.1	$\pm 3.0$	
	P-LCC $\times 10^9/L$ **	15	$\pm 15$	70	$\pm 25$	115	$\pm 35$	
	P-LCR %**	34.0	$\pm 10.0$	28.5	$\pm 10.0$	23.5	$\pm 10.0$	
BC-5300Vet,BC-5100Vet QC Mode	WBC $\times 10^9/L$	3.25	$\pm 0.50$	7.80	$\pm 1.00$	17.80	$\pm 2.50$	
	Neu# $\times 10^9/L$	1.64	$\pm 0.30$	4.56	$\pm 0.63$	11.71	$\pm 1.43$	
	Lym# $\times 10^9/L$	1.25	$\pm 0.30$	2.22	$\pm 0.63$	3.51	$\pm 1.43$	
	Mon# $\times 10^9/L$	0.16	$\pm 0.13$	0.31	$\pm 0.24$	0.71	$\pm 0.54$	
	Eos# $\times 10^9/L$	0.20	$\pm 0.17$	0.70	$\pm 0.55$	1.87	$\pm 1.43$	
	Neu%	50.5	$\pm 9.0$	58.5	$\pm 8.0$	65.8	$\pm 8.0$	
	Lym%	38.5	$\pm 9.0$	28.5	$\pm 8.0$	19.7	$\pm 8.0$	
	Mon%	5.0	$\pm 4.0$	4.0	$\pm 3.0$	4.0	$\pm 3.0$	
	Eos%	6.0	$\pm 5.0$	9.0	$\pm 7.0$	10.5	$\pm 8.0$	
	RBC $\times 10^{12}/L$	2.03	$\pm 0.18$	4.17	$\pm 0.24$	5.00	$\pm 0.30$	
	HGB g/L	56	$\pm 4$	126	$\pm 6$	161	$\pm 8$	
	HCT %	18.3	$\pm 1.5$	42.0	$\pm 2.0$	54.3	$\pm 2.4$	
	MCV fL	90.0	$\pm 5.0$	100.8	$\pm 5.0$	108.5	$\pm 5.0$	
	MCH pg	27.6	$\pm 2.5$	30.2	$\pm 2.5$	32.2	$\pm 2.5$	
	MCHC g/L	307	$\pm 30$	300	$\pm 30$	297	$\pm 30$	
	RDW-CV %	14.0	$\pm 3.0$	13.5	$\pm 3.0$	13.0	$\pm 3.0$	
	RDW-SD fL	56.0	$\pm 8.0$	59.0	$\pm 8.0$	60.8	$\pm 8.0$	
	PLT $\times 10^9/L$	54	$\pm 20$	236	$\pm 40$	475	$\pm 60$	
	MPV fL	9.2	$\pm 3.0$	8.6	$\pm 3.0$	8.1	$\pm 3.0$	
	PCT %*	0.050	$\pm 0.050$	0.200	$\pm 0.100$	0.385	$\pm 0.200$	
	PDW*	16.8	$\pm 3.0$	16.5	$\pm 3.0$	16.0	$\pm 3.0$	

\* For Research Use Only

\*\* These parameters are not provided on BC-5000/BC-5000 Vet analyzers

Before using, refer to the instruction sheet for mixing directions.

All brands and products are trademarks or registered trademarks of their respective companies.



Mindray

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China

Tel: +86 755 81888998

Fax: +86 755 26582680

EC REP

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726