


BC-RET

HEMATOLOGY CONTROLS

CONTROL

LOT ME0319A

 **2019-05-10**

Instrument	Parameter	Low		Normal		High	
		LOT	ME0319A-1	LOT	ME0319A-2	LOT	ME0319A-3
MINDRAY	Ret# $10^{12}/L$	0.0682	± 0.0546	0.2041	± 0.0816	0.1906	± 0.0572
BC-6800, BC-6600	Ret% %	1.21	± 0.97	4.39	± 1.76	7.94	± 2.38
	IRF %	7.7	± 15.0	6.9	± 13.0	6.2	± 25.0
QC Mode	LFR %	92.3	± 15.0	93.1	± 10.0	93.8	± 25.0
	MFR %	6.3	± 12.0	5.6	± 10.0	4.7	± 22.0
	HFR %	1.4	± 3.0	1.3	± 3.0	1.5	± 3.0
	RBC-O* $10^{12}/L$	5.65	± 0.35	4.65	± 0.30	2.40	± 0.20

Before using, refer to the instruction sheet for mixing directions.

*Research use only.

Instrument	Parameter	Low		Normal		High	
		LOT	ME0319A-1	LOT	ME0319A-2	LOT	ME0319A-3
MINDRAY	Ret# $10^{12}/L$	0.0735	± 0.0230	0.2208	± 0.0990	0.1974	± 0.0770
BC-6200,	Ret% %	1.26	± 0.40	4.59	± 2.00	7.96	± 3.00
BC-6000Plus,	IRF %	11.8	± 15.0	10.5	± 13.0	8.9	± 25.0
BC-6100Plus	LFR %	88.2	± 15.0	89.5	± 10.0	91.1	± 25.0
QC Mode	MFR %	9.1	± 12.0	7.4	± 10.0	5.2	± 22.0
	HFR %	2.7	± 3.0	3.1	± 3.0	3.7	± 3.0
	RHE pg	30.1	± 10.0	30.3	± 10.0	24.5	± 10.0

Before using, refer to the instruction sheet for mixing directions.

Instrument	Parameter	Low		Normal		High	
		LOT	ME0319A-1	LOT	ME0319A-2	LOT	ME0319A-3
MINDRAY	Ret# $10^{12}/L$	0.0827	± 0.0230	0.2181	± 0.0990	0.1944	± 0.0770
BC-6800Plus,	Ret% %	1.45	± 0.40	4.61	± 2.00	7.87	± 3.00
BC-6700Plus,	IRF %	10.7	± 15.0	10.1	± 13.0	9.0	± 25.0
BC-6600Plus	LFR %	89.3	± 15.0	89.9	± 10.0	91.0	± 25.0
QC Mode	MFR %	7.8	± 12.0	7.0	± 10.0	5.2	± 22.0
	HFR %	2.9	± 3.0	3.1	± 3.0	3.8	± 3.0
	RHE pg	29.4	± 10.0	29.6	± 10.0	24.2	± 10.0

Before using, refer to the instruction sheet for mixing directions.



Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, ShenZhen 518057, P.R.China

Tel: +86 755 26582479 26582888

Fax: +86 755 26582934 26582500

EC REP

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80 20537 Hamburg, Germany

Tel: 0049-40-2513175

Fax: 0049-40-255726



P/N:046-003620-00(3.0)