


Research use only

# BC-RET

## HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

**LOT** ME0919A  
 2019-11-10

Instrument	Parameter	Low		Normal		High	
		<b>LOT</b>	ME0919A-1	<b>LOT</b>	ME0919A-2	<b>LOT</b>	ME0919A-3
MINDRAY BC-6800, BC-6600	PLT-O 10 <sup>9</sup> /L		125 ± 20		274 ± 40		516 ± 60
QC Mode							

Instrument	Parameter	Low		Normal		High	
		<b>LOT</b>	ME0919A-1	<b>LOT</b>	ME0919A-2	<b>LOT</b>	ME0919A-3
MINDRAY BC-6200, BC-6000Plus, BC-6100Plus	RBC-O 10 <sup>12</sup> /L		5.82 ± 0.35		4.68 ± 0.30		2.36 ± 0.20
	PLT-O 10 <sup>9</sup> /L		128 ± 20		292 ± 40		524 ± 60
QC Mode							

Instrument	Parameter	Low		Normal		High	
		<b>LOT</b>	ME0919A-1	<b>LOT</b>	ME0919A-2	<b>LOT</b>	ME0919A-3
MINDRAY BC-6800Plus, BC-6700Plus, BC-6600Plus	RBC-O 10 <sup>12</sup> /L		5.70 ± 0.35		4.64 ± 0.30		2.40 ± 0.20
	PLT-O 10 <sup>9</sup> /L		121 ± 20		291 ± 40		537 ± 60
QC Mode							

Research use only