

S30 CALIBRATOR

FOR MINDRAY HEMATOLOGY ANALYZERS

LOT

BC10519C-1



2019-09-18

Calibration Values

Analyzer	Parameter	Values	Range
BC-3600, BC-3300	WBC $\times 10^9/L$	9.2	± 0.3
	RBC $\times 10^{12}/L$	4.30	± 0.10
	HGB g/L	132	± 2
	MCV fL	88.9	± 2.0
	PLT $\times 10^9/L$	256	± 15
BC-3200, BC-3000CT	WBC $\times 10^9/L$	9.0	± 0.3
	RBC $\times 10^{12}/L$	4.42	± 0.10
	HGB g/L	133	± 2
	MCV fL	89.2	± 2.0
	PLT $\times 10^9/L$	240	± 15
BC-2900, BC-1800, BC-3000 Plus	WBC $\times 10^9/L$	9.1	± 0.3
	RBC $\times 10^{12}/L$	4.41	± 0.10
	HGB g/L	135	± 2
	MCV fL	89.3	± 2.0
	PLT $\times 10^9/L$	243	± 15
BC-2800, BC-2800Vet, BC-2600, BC-2600Vet	WBC $\times 10^9/L$	9.2	± 0.3
	RBC $\times 10^{12}/L$	4.28	± 0.10
	HGB g/L	134	± 2
	MCV fL	90.0	± 2.0
	PLT $\times 10^9/L$	254	± 15
BC-2300, BC-2100	WBC $\times 10^9/L$	9.2	± 0.3
	RBC $\times 10^{12}/L$	4.09	± 0.10
	HGB g/L	132	± 2
	MCV fL	88.6	± 2.0
	PLT $\times 10^9/L$	246	± 15
BC-20s, BC-21s BC-30s, BC-31s BC-30, BC-31, BC-30e	WBC $\times 10^9/L$	9.4	± 0.3
	RBC $\times 10^{12}/L$	4.17	± 0.10
	HGB g/L	135	± 2
	MCV fL	89.9	± 2.0
	PLT $\times 10^9/L$	239	± 15
BC-20, BC-21 BC-10, BC-11, BC-10e	WBC $\times 10^9/L$	9.2	± 0.3
	RBC $\times 10^{12}/L$	4.22	± 0.10
	HGB g/L	134	± 2
	MCV fL	90.5	± 2.0
	PLT $\times 10^9/L$	250	± 15